

## INFLUVAC® TETRA 2018 FLAG LABELS

Dear User

Please note that the additional flag label which is designed to be easily removed by peeling off and tearing along a serrated line on the label, is not in most cases completely adhered to the syringe barrel.

We apologise for any inconvenience that this may cause.

Yours sincerely,

Lloyd Price  
Country Manager  
MYLAN NEW ZEALAND

